

Accu-Chek Performa meter operator's manual.pdf - Adobe Acrobat Pro

File Edit View Window Help


Create

9 (11 of 192) 150%

Tools Comment Share

Coding the Meter

9



1. Make sure the meter is off.
2. Turn the meter over.
3. Remove the old code chip (if there is one in the meter) and discard it.

Change the code chip every time a new box of test strips is opened!

Accu-Chek Performa meter operator's manual.pdf - Adobe Acrobat Pro


File Edit View Window Help

Create

10 (12 of 192) 158%

Tools Comment Share

10



4. Turn the code chip over so the code number faces down. Push the code chip into the code chip slot until it stops.
5. Leave the code chip in the meter until a new box of test strips is opened.

Notes:

- Do not force the code chip into the meter. It is designed to go into the meter only one way.
- If code "--" is displayed, insert a code chip into the meter.